

EPJ Web of Conferences **88**, 01013 (2015)

DOI: 10.1051/epjconf/20158801013

© Owned by the authors, published by EDP Sciences - SIF, 2015

Extracting information from partially depleted Si detectors with digital sampling electronics

G. PASTORE^{1,2}, G. PASQUALI^{1,2}, N. LE NEINDRE³, G. ADEMARD⁴,
S. BARLINI^{1,2}, M. BINI^{1,2}, E. BONNET⁵, B. BORDERIE⁴,
R. BOUGAULT³, G. CASINI², A. CHBIHI⁵, M. CINAUSERO⁶,
J.A. DUEÑAS⁷, P. EDELBRUCK⁴, J.D. FRANKLAND⁵, F. GRAMEGNA⁶,
D. GRUYER⁵, A. KORDYASZ⁸, T. KOZIK⁹, O. LOPEZ³, T. MARCHI⁶,
L. MORELLI¹⁰, A. OLM², A. ORDINE¹¹, M. PÂRLOG^{3,12}, S. PIANTELLI²,
G. POGGI^{1,2}, M.-F. RIVET⁴, E. ROSATO^{11,13}, F. SALOMON⁴,
G. SPADACCINI^{11,13}, A.A. STEFANINI^{1,2}, S. VALDRÉ^{1,2}, E. VIENT³,
T. TWARÓG⁹, R. ALBA¹⁴, C. MAIOLINO¹⁴ and D. SANTONOCITO¹⁴

¹ Dipartimento di Fisica, Università di Firenze, Firenze, Italy

² INFN, Sezione di Firenze, Firenze, Italy

³ LPC, IN2P3-CNRS, ENSICAEN and Université de Caen, France

⁴ IPN, CNRS/IN2P3, Université Paris Sud 11, Paris, France

⁵GANIL, CEA/DSM-CNRS/IN2P3, Caen, France

⁶INFN-LNL Legnaro (PD), Italy

⁷Dep. de Fisica Aplicada, FCCEE Univ. de Huelva, Huelva, Spain

⁸Heavy Ion Laboratory, University of Warsaw, Warsaw, Poland

⁹Jagiellonian Univ., Institute of Nuclear Physics IFJ-PAN, Kraków, Poland

¹⁰INFN, Sezione di Bologna and Università di Bologna, Italy

¹¹INFN, Sezione di Napoli, Napoli, Italy

¹²NIPNE-HH, Măgurele-Bucharest, Romania

¹³Dipartimento di Fisica, Università di Napoli “Federico II”, Napoli, Italy

¹⁴INFN-LNS, Catania, Italy

Abstract

A study of the identification properties and of the energy response of a Si-Si-CsI(Tl) ΔE - E telescope exploiting a partially depleted second Si stage has been performed. Five different bias voltages have been applied to the second stage of the telescope, one corresponding to full depletion, the others associated with a depleted layer ranging from 60% to 90% of the detector thickness. Fragment identification has been obtained using either the ΔE - E technique or the Pulse Shape Analysis (PSA). Charge collection efficiency has been evaluated. The ΔE - E performance is not affected by incomplete depletion. Isotopic separation capability improves at lower bias voltages with respect to full depletion, though charge identification thresholds increase.

In recent years, intensive experimental work of the FAZIA Collaboration [1,2] has been devoted to improving the nuclear fragment identification based on Pulse Shape Analysis. Since an improvement of PSA isotopic identification capabilities in partially depleted Si detectors had been observed in preliminary tests, a systematic study of PS identification capabilities of underdepleted Si detectors has been undertaken [3].

The second stage of a standard FAZIA telescope, Si2, has been biased at five different voltages (see tab.1). In fact, the PSA performance can be better studied on Si2 since charge and mass of the fragments stopped in Si2 can be obtained from the ΔE - E technique applied to fragments that punch through the first stage of the telescope.

Table 1: Main features of the detector at different bias voltages.

Voltage (V)	105	130	200	235	290
Depletion depth (μm)	310	340	420	460	510
Undepleted layer (μm)	200	170	90	50	0

The data were collected in Catania at Laboratori Nazionali del Sud (LNS) of INFN. The beam was ^{84}Kr at 35A MeV impinging on ^{112}Sn and ^{197}Au targets. A telescope composed of Si 300 μm -Si 500 μm -CsI(Tl) 10 cm, was mounted in the Ciclope scattering chamber, at a distance from target of 50 cm and at 8° polar angle with respect to the beam line, covering about 2° in polar angle. According to the usual FAZIA recipe for optimizing the PSA performance, both silicon detectors were mounted with the ohmic (low field) side facing the target [4]. The Si detectors were of the neutron transmutation doped type for a better doping uniformity and cut at a random direction to avoid channeling effects [5]. They were connected to PACI preamplifiers placed under vacuum very close to the detectors. Signals were

then brought outside vacuum and connected to custom made digitizers [2]. For Si2, since charge collection times in an underdepleted detector can be longer than at full depletion, the length of acquired signals has been set to $70\ \mu\text{s}$ (maximum length allowed by the FEE signal memory).

An accurate study of the charge collection efficiency showed for the two lowest biases a dependence on the range of the stopped fragments: the higher the range of the projectile, the better the collection efficiency of the detector (about 98% for particles that punch through Si2). We find that it is possible to reproduce the behaviour of the collection efficiency with a simple formula exploiting the variation of the stopping power along the particle track. Details can be found in [3].




Figure 1: ΔE - E and PSA correlations obtained at the highest (top panels) and lowest (bottom panels) bias voltages.

Coming to the ΔE - E identification of fragments, the left panels of fig.1 show the correlations obtained at the highest (top panel) and lowest (bottom panel) bias voltages applied to Si2. No significant degradation of the isotopic separation can be spotted in the figure though at 105 V, Si2 is only depleted by 60%. The right panels of fig.1 show the PSA correlations (“Energy in Si2 vs maximum of current signal”) obtained at the highest (top panel) and lowest (bottom panel) bias voltages. The comparison evidences a sizable improvement of the identification capabilities of the detector when underbiased.

The energy threshold for discrimination can be quantitatively evaluated for each element by calculating a Figure of Merit as a function of fragment energy [2, 3]. Figure 2 presents the energy threshold for identification both in mass (left panel) and in charge (right panel) as a function of the atomic number Z . One can observe that the lower the applied voltage, the higher the energy threshold for charge identification for a given Z and the larger the maximum Z for which isotopic identification is achieved. A doping uniformity of about 1% FWHM, or less, would be needed to discriminate isotopes via PSA in a totally depleted detector [6]. The doping uniformity of the detector under test is about 6%: in fact only when underbiased it has been found capable of identifying isotopes, even though in a reduced energy domain.




Figure 2: The energy thresholds respectively for isotopic identification on the left, for charge identification on the right.

As a conclusion, we observed a relatively high charge collection efficiency (more than about 90%) even when only 60% of the detector thickness is depleted, identification via $\Delta E-E$ is not affected by underbias, isotopic identification via PSA improves with underbias though one gets higher energy thresholds for charge identification with respect to full depletion.

References

- [1] FAZIA Collaboration, <http://fazia2.in2p3.fr/spip>
- [2] R. Bougault et al., *Eur. Phys. J. A*, **50** (2014) 47.
- [3] G.Pasquali et al., *Eur. Phys. J. A*, **50** (2014) 86.
- [4] N.Le Neindre et al., *Nucl. Instrum. Methods A*, **701** (2013) 145.
- [5] S.Carboni et al., *Nucl. Instrum. Methods A*, **664** (2012) 251.
- [6] L.Bardelli et al., *Nucl. Instrum. Methods A*, **654** (2011) 272.